

# **Mascarilla FFP2 NR**



## UNIVERSAL CERTIFICATION and SURVEILLANCE SERVICES TRADE CO. Necip Fazil Bulvari Keyap Sitesi E2 Blok No:44/84 Yukari Dudullu Umraniye, Istanbul / TURKEY

#### TEST REPORT

Report Date: 26.05.2020 Report Number: 05-2020-T-095

#### CLIENT and SAMPLE INFORMATION

TEST OWNER	Changzou 1	Changzou Ruida Medical Technology Co. Ltd.				
ADDRESS	No: 88 Mahang Middle Road, Hutang Town, Wujin District, Changzhou City, Jiangsu Province China					
SAMPLE DESCRIPTION	Vertical flat	Vertical flat fold (See end of test report for sample photo)				
BRAND NAME - MODEL	PROMASK	/ KN 95				
TESTING STANDARD	EN 149+A1	EN 149+A1:2009				
CASE NUMBER	CE-PPE-2078					
SAMPLE RECEIVE DATE	11.05.2020	11.05.2020 TESTING START DATE 12.0		12.05.2020		
DISINFECTION INSTRUCTION  If applicable	Not given, s	ingle use only				
NUMBER OF SAMPLES	50	SAMPLE I	Ds:	1 – 46		
AS RECEIVED SAMPLE NO	26-46					
	Simulated wearing treatment		1-2-3-4-5-6-7-8-9 (As Received)			
CONDITIONING SAMPLE NO	Temperature conditioning		10-11-12-13-14-15 (Sample after test of Mechanical Strength)			
			16-	17-18-19-20-21-22-23	3-24-25 (As Received)	
	Mechanical	strength	10-	11-12-13-14-15 (As R	Received)	

The results given in this test report belongs to the samples tested. The report content cannot be recreated partially without the written consent of UNIVERSAL CERTIFICATION.

UNIVERSAL SERTIFICASYON

F SYLTHING ATI

Neclo Faril Rulvar, Kys.

Yukan Dudha, Di Hanya, STANBUL

Telefon: 0216 455 00.00 5 0216 #55 00.00

Saripaz Stanb KAC MAZ

Director

Page 1 / 12

UNIVERSAL SERTIFIKASYON VE GÖZETİM HİZM. TİC. LTD. ŞTİ. Keyap Ticaset Merkezi, Necip Fani Bulvan, E2 Blok, No:44/84 Y, Dodullu - Ümraniye - İSTANBUL T1-90 216 455 80 80 F-90 216 455 80 F-90 216 455 80 80 F-90 216 455 80 80 F-90 216 455 80 F-90 216 455 80 F-90 216 455 80 F-90 216





#### 1. REPORT SUMMARY

TEST STANDARD	TEST NAME	RESULT	EVALUATION
EN 149:2001 + A1:2009 clause 8.5 EN 13274-1:2001	Total Inward Leakage Testing	Pass	FFP2
EN 149:2001 + A1:2009 clause 8.11 EN 13274-7:2019	Penetration of Filter Material	Pass	FFP2
EN 149:2001 + A1:2009 clause 8.6 EN 13274-4:2001	Flammability Testing	Pass	See result
EN 149:2001 + A1:2009 clause 8.7 EN 13274-6:2001	Carbon Dioxide Content of The Inhalation Air Testing	Pass	See result
EN 149:2001 + A1:2009 clause 8.9	Breathing Inhalation Resistance-30 l/min	Pass	See results
EN 13274-3:2001	Breathing Inhalation Resistance-95 l/min	Pass	See results
EN 149:2001 + A1:2009 clause 8.9 EN 13274-3:2001	Exhalation Resistance, flow rate 160 l/min	Pass	See result



Page 2 / 12

UNIVERSAL SERTIFIKASYON VE GÖZETİM HİZM. TİC. LID. ŞTİ. Keyap Ticaret Merkezi, Necip Faril Bulvan, E2 Blok, No.94/84 Y. Dudullu - Ümraniye - İSTANBUL T.+90 216 455 80 80 F.+90 216 455 80 08 info@universalcert.com





#### 2. TEST RESULTS and EVALUATION

## 7.4 PACKAGING (EN 149:2001 + A1:2009 clause 8.2)

Test Method: Clause 8.2-Visual inspection

REQUIREMENT	RESULTS	COMMENT
Particle filtering half masks shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use.	Pass	The masks were packaged in sealed plastic bags, in larger plastic bags inside a large cardboard box that gave some protection against mechanical damage or contamination before use

Lab A

## 7.5 MATERIAL (EN 149:2001 + A1:2009 clause 8.2, 8.3.1, 8.3.2)

Test Method: Clause 8.2-Visual inspection

Clause 8.3.1-Simulated wearing treatment

A breathing machine is adjusted to 25 cycles/min and 2,0 l/stroke. The particle filtering half mask was mounted on a Sheffield dummy head.

For testing, a saturator is incorporated in the exhalation line between the breathing machine and the dummy head, the saturator being set at a temperature in excess of 37 °C to allow for the cooling of the air before it reaches the mouth of the dummy head.

The air has been saturated at  $(37 \pm 2)$  °C at the mouth of the dummy head

Clause 8.3.2-Temperature conditioning

The ambient temperature for testing has been between 16 °C and 32 °C and the temperature limits has been subject to an accuracy of  $\pm 1$  °C.

- a) for 24 h to a dry atmosphere of  $(70 \pm 3)$  °C;
- b) for 24 h to a temperature of  $(-30 \pm 3)$  °C; and allow to return to room temperature for at least 4 h between exposures and prior to subsequent testing. The conditioning has been carried out in a manner which ensures that no thermal shock occurs.

REQUIREMENT	RESULTS	COMMENT
Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used.	Pass	The materials used were able to withstand handling and wear during the limited laboratory testing carried out.
Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer.	Pass	It was not constitute a hazard or nuisance for the wearer.
After undergoing the conditioning described in 8.3.1 none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps.	Pass	None of the specimens conditioned suffered mechanical failure.
When conditioned in accordance with 8.3.1 and 8.3.2 the particle filtering half mask shall not collapse.	Pass	None of the specimens had not collapse after conditioning.

Lab B

de

Page 3 / 12





# 7.6 CLEANING AND DISINFECTING (EN 149:2001 + A1:2009 clause 8.4, 8.5, 8.11)

Test Method: Described in Clause 8.4, 8.5 and 8.11

REQUIREMENT	RESULTS	COMMENT
If the particle filtering half mask is designed to be re- usable, the materials used shall withstand the cleaning and disinfecting agents and procedures to be specified by the manufacturer. With reference to 7.9.2, after cleaning and disinfecting the re-usable particle filtering half mask shall satisfy the penetration requirement of the relevant class.	N/A	This article is not applicable for tested protective mask which is single use disposable mask.

# 7.7 PRACTICAL PERFORMANCE (EN 149:2001 + A1:2009 clause 8.4)

Test Method: Described in Clause 8.4

REQUIREMENT	RESULTS	COMMENT
The particle filtering half mask shall undergo practical performance tests under realistic conditions. These general tests serve the purpose of checking the equipment for imperfections that can not be determined by the tests described elsewhere in this standard.	No imperfections	Detail refer to Annex I

#### Annex I-Test Result:

Number of sample: 29 (A.R), 30 (A.R)

Assessed elements	Positive Assessment	Negative Assessment	Requirements in accordance with EN 149:2001+A1:2009	Assessment of Test Result Conformity / Nonconformity
The face piece fitting Head harness comfort Security of fastenings Speech clearness Field of vision Materials compatibility with skin	2 2 2 2 2 2 2 2	0 0 0 0 0	Filtering half masks should not have imperfections related to wearer's acceptance	Filtering half masks fulfil requirements of the standard EN 149:2001 + A1:2009 given in 7.7

Lab B



Page 4 / 12

UNIVERSAL SERTIFIKASYON VE GÖZETİM HİZM. TİC, LITD, ŞTİ. Keyup Ticaret Merkezi, Necip Fazal Bulvarı, E2 Blok, No:44/84 Y. Dudullu - Ürmneiye - İSTANBUL T:+90 216 455 80 80 F:490 216 455 80 08 info@universalcet.com





## 7.8 FINISH OF PARTS (EN 149:2001 + A1:2009 clause 8.2)

Test Method: Described in Clause 8.2

REQUIREMENT	RESULTS	COMMENT
Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs.	Pass	None of the specimens used in laboratory testing showed evidence of sharp edges or burrs.

Lab A

# 7.9.1 TOTAL INWARD LEAKAGE (EN 149:2001 + A1:2009 clause 8.5)

#### Test Method: Described in Clause 8.5

REQUIREMENT	RESULTS	COMMENT	
The total inward leakage consists of three components: face seal leakage, exhalation value leakage (if exhalation value fitted) and filter penetration. For particle filtering half masks fitted in accordance with the manufacturer's information, at least 46 out of the 50 individual results shall be not greater than: 25 % for FFP1, 11 % for FFP2, 5 % for FFP3 and in addition at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall not be greater than: 22 % for FFP1, 8 % for FFP2 2 % for FFP3	Pass	Classified as FFP2  Detail refer to Annex II	

## Annex II-Test Result:

The test results obtained are given in the tables as follows

Test Subject	No of sample	Cond.	1. Walk (%)	Head side/ side (%)	Head up/down (%)	Talk (%)	2. Walk (%)	Mean (%)
1	31	A.R.	6,4	5 47	4,15	6,13	6,79	5,87
2	32	A.R.	6,53	4,45	6,17	7,29	7,45	6,38
3	33	A.R.	7,36	6,92	7,1	6,19	7,12	6,94
4	34	A.R.	6,33	5,24	7,32	6,26	7,4	6,51
5	35	A.R.	6,81	6,41	7,5	6,12	8,22	7,01
6	16	T.C.	5,63	7,45	6,75	7,21	6,2	6,65
7	17	T.C.	6,85	7,8	7,61	6,29	7,32	7,17
8	18	T.C.	7,69	7,9	8,96	8,28	8,19	8,2
9	19	T.C.	7,58	8,19	7,37	7,6	9,2	7,99
10	20	T.C.	7,57	7.41	7,42	7,76	8,87	7,81

Lab B



Page 5 / 12

UNIVERSAL SERTIFIKASYON VE GÖZETIM HIZM. TIC. LTD, ŞTİ, Keyap Ticaret Merkezi, Necip Fazd Bulvum, E2 Bick, Nov4484 Y. Dochulu - Ümraniye - İSTANBUL T. 9/0 216 455 80 80 F; 9/0 216 455 80 80 Hinfo@universalcent.com





# 7.9.2 PENETRATION OF FILTER MATERIAL (EN 149:2001 + A1:2009 clause 8.11)

Test Method: Described in Clause 8.11

REQUIREMENT			RESULTS	COMMENT
Classification Max penetration of test aerosol		on of test aerosol		
	NaCl test 95 l/min %max	Paraffin oil test 95 l/min %max	Pass	Detail refer to Annex IIIA and IIIB
FFP1	20	20		Detail refer to funda first and first
FFP2	6	6		
FFP3	1	1		

## Annex IIIA-Test Result:

The test results obtained are given in the tables as follows:

No. of Sample	Condition	Penetration of Sodium Chloride in accordance with EN 13274-7:2019 [%] Flow rate 95 l/min	Requirements in accordance with EN 149:2001+A1:2009	Assessment of Test Result Conformity / Nonconformity
36		2,65		Passed
37	As received	4,03		1
38		5,07	FFP1 ≤ 20 %	Filtering half masks fulfil
1	Cimulated manning	2,70	FFF1 \sec. 20 %	the requirements of the
2	Simulated wearing treatment	4,06	FFP2 ≤ 6 %	standard EN
3	treatment	5,10	1112 20 70	149:2001+A1:2009 given
10	Mechanical strength +	2,68	FFP3 ≤ 1 %	in 7.9.2 in range of the
11	Temperature	4,06	11172170	first and second
12	conditioned	5,09		protection classes (FFP1, FFP2)

## Annex IIIB-Test Result:

The test results obtained are given in the tables as follows;

No. of Sample	Condition	Penetration of Paraffin Oil Mist in accordance with EN 13274-7:2019 [%] Flow rate 95 l/min	Requirements in accordance with EN 149:2001+A1:2009	Assessment of Test Result Conformity / Nonconformity	
39	10 10 10	3,97		Passed	
40	As received	3,18		C.F. ARRESTA	
41		4,58			
4	Classification and a second	4,01	FFP1 ≤ 20 %	Filtering half masks	
5	Simulated wearing treatment	3,20		fulfil the requirements	
6	treatment	4,61	FFP2 ≤ 6 %	of the standard EN	
13		3,99		149:2001+A1:2009	
14	Mechanical strength +	3,21	FFP3 ≤ 1 %	given in 7.9.2 in range	
15	Temperature conditioned	4,63		of the first and second protection classes (FFP1, FFP2)	

Lab A+B



Page 6 / 12

UNIVERSAL SERTIFIKASYON VE GÖZETIM HİZM. TİC. ETD. ŞEL Keyap Ticaret Meskezi, Nacip Fazil Bulvan, E.2 Blok, Noc44/64 Y. Dudullu - Ümraniye - İSTANBUL T:+90 216 455 80 80 F:+90 216 455 80 08 info@universalcort.com





# 7.10 COMPATIBILITY WITH SKIN (EN 149:2001 + A1:2009 clause 8.4, 8.5)

Test Method: Described in Clause 8.4 and 8.5.

REQUIREMENT	RESULTS	COMMENT
Materials that may come into contact with the wearer's skin shall not be known to be likely to cause irritation or any other adverse effect to health.	Pass	No irritation or any other adverse effect to health or sensitivity reported by the subjects.

Lab B

## 7.11 FLAMMABILITY (EN 149:2001 + A1:2009 clause 8.6)

#### Test Method: Described in Clause 8.6

REQUIREMENT	RESULTS	COMMENT	
The material used shall not present a danger for the wearer and shall not be of highly flammable nature. When tested, the particle filtering half mask shall not burn or not to continue to burn 5s after removal from the flame.	Pass	Detail refer to Annex IV	

#### Annex IV-Test Result:

The test results obtained are given in the tables as follows:

No. of Sample	Condition	Visual inspection	Requirements in accordance with EN 149:2001+A1:2009	Assessment of Test Result Conformity / Nonconformity
45	As received	As regained 2,40		Passed
46	Asteceived	2,08	Filtering half mask	
21		2,19	shall not burn or not continue to burn for	Filtering half masks fulfil requirements of
22	Temperature conditioned	2,14	more than 5 s after removal from the flame	the standard EN 149:2001 + A1:2009 given in 7.11

Lab B



Page 7/12

UNIVERSAL, SERTIFIKASYON VE GÖZETIM HIZM. TIC, LTD. ŞTİ. Keyap Ticarer Merkezi, Nəcip Fazıl Bulvan, E2 Blok, No:34/84 Y, Dudullu - Üzmaniye - İSTANBUL T:+90 216 455 80 80 P:+90 216 455 80 08 info@universaleert com





# 7.12 CARBON DIOXIDE CONTENT OF THE INHALATION AIR (EN 149:2001 + A1:2009 clause 8.7)

#### Test Method: Described in Clause 8.7

REQUIREMENT	RESULTS	COMMENT	
The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1,0 % (by volume)	Pass	Detail refer to Annex V	

#### Annex V-Test Result:

The test results obtained are given in the tables as follows:

No. of Sample	Condition	CO <sub>2</sub> content of the inhalation air [%] by volume	An average CO <sub>2</sub> content of the inhalation air [%] by volume	Requirements in accordance with EN 149:2001+A1:2009	Assessment of Test Result Conformity / Nonconformity
26		0,49		CO2 content of the	Passed
27	As received	0,64	0,57	inhalation air shall	Filtering half masks fulfil
28		0,59	not exceed an average of 1,0% by volume	requirements of the standard EN 149:2001 + A1:2009 given in 7.12	

Lab B

## 7.13 HEAD HARNESS (EN 149:2001 + A1:2009 clause 8.4, 8.5)

# Test Method: Described in Clause 8.4, 8.5

REQUIREMENT	RESULTS	COMMENT
The head harness shall be designed so that the particle filtering half-mask can be donned and removed easily.	Pass	No problem with the head harness reported by the wearers during the practical performance test.
The head harness shall be adjustable or self-adjusting and shall be sufficiently robust to hold the particle filtering half mask firmly in position and capable of maintaining total inward leakage requirements for the device.	Pass	No problem with the head harness reported by the wearers during the practical performance test.

Lab B



Page 8 / 12

UNIVERSAL SERTIFIKASYON VE GÖZETIM HİZM. TİC. LTD. ŞTİ. Keyap Ticarer Medsezi, Necip Fazıl Bulvarı, E2 Biok, Noci4/84 Y, Dudullu - Ümminiye - İSTANBUL T.+90 216 455 80 80 F;+90 216 455 80 08 info@universalcert.com





# 7.14 FIELD OF VISION (EN 149:2001 + A1:2009 clause 8.4)

Test Method: Described in Clause 8.4

REQUIREMENT	RESULTS	COMMENT
The field of vision is acceptable if determined so in practical performance tests.	Pass	There were no adverse comments following practical performance tests.

Lab B

## 7.15 EXHALATION VALVE (EN 149:2001 + A1:2009 clause 8.2, 8.3.4, 8.8, 8.9.1)

Test Method: Clause 8.2, 8.3.4, 8.8, 8.9.1

REQUIREMENT	RESULTS	COMMENT
A particle filtering half mask may have one or more exhalation valve(s), which shall function correctly in all orientations.	N/A	No exhalation valve in tested samples.
If an exhalation valve is provided it shall be protected against or be resistant to dirt and mechanical damage and may be shrouded or may include any other device that may be necessary for the particle filtering half mask to comply with 7.9	N/A	No exhalation valve in tested samples.
Exhalation valve(s), if fitted, shall continue to operate correctly after a continuous exhalation flow of 300 l/min over a period of 30s.	N/A	No exhalation valve in tested samples.
When the exhalation valve housing is attached to the face blank, it shall withstand axially a tensile force of 10N applied for 10s.	N/A	No exhalation valve in tested samples.

Lab -



Page 9 / 12

UNIVERSAL SERTIFIKASYON VE GÖZETİM HİZM. TIC. LTD. ŞTİ. Keyap Ticaret Merkezi, Necip Fanil Bulvarı, E2 Blok, Neci-4/84 Y, Dudullu - Ümraniye - ESTANBUL, T.+90 216 455 80 80 F.+90 216 455 80 88 info@miversalcert.com





## 7.16 BREATHING RESISTANCE (EN 149:2001 + A1:2009 clause 8.9)

## Test Method: Described in Clause 8.9

	REQU	IREMENT		RESULTS	COMMENT
Classification [	Max per	mitted resistance	e (mbar)		Classified as FFP3
	Inhal	lation	Exhalation		
	30 l/min	95 l/min	160 l/min	Pass	Detail refer to Annex VIA-VIB
FFP1	0.6	2.1	3.0		Detail ferei to runiex VIA-VIB
FFP2	0.7	2.4	3.0		
FFP3	1.0	3.0	3.0		

#### Annex VIA-Test Result:

The test results obtained are given in the tables as follows;

Inhal	ation	Res	istanc	E
-------	-------	-----	--------	---

No. of	Condition		Inhalation Resistance (mbar)					
Sample		Flow rate 30 l/min	Requirements in accordance with EN 149:2001+A1:2009	Flow rate 95 I/min	Requirements in accordance with EN 149:2001+A1:2009	Assessment of Test Result Conformity / Nonconformity		
42		0,36		1,21		Passed		
43	As received	0,38		1,25		Passed		
44		0,37	FFP1 ≤ 0,60	1,23	FFP1 ≤ 2,10	Passed		
7	Simulated	0,38		1,23	1111 52,10	Passed		
8	wearing	0,40	FFP2 ≤ 0,70	1,27	FFP2 ≤ 2,40	Passed		
9	treatment	0,39		1,26		Passed		
23	T	0,37	FFP3 ≤ 1,0	1,25	FFP3 ≤ 3,00	Passed		
24	Temperature conditioned	0,38		1,28		Passed		
25	conditioned	0,38		1,29		Passed		

Exhalation	Resistance
------------	------------

No. of Sample	Condition	Flow rate	Facing directly	Facing vertically upwards	Facing vertically downwards	Lyin g on the left side	Lying on the right side	Requirements in accordance with EN 149:2001+A1:2009	Assessment of Test Result Conformity / Nonconfor mity	
42	As received			2,18	2,20	2,21	2,20	2,21		Passed
43			2,23	2,25	2,25	2,27	2,29		Passed	
44			2,24	2,26	2,27	2,30	2,31	FFP1 ≤ 3,0	Passed	
7	Simulated	lated	2,20	2,22	2,25	2,28	2.26	1111 55,0	Passed	
8	wearing	1601/min	2,25	2,30	2,32	2,33	2,30	FFP2 ≤ 3.0	Passed	
9	treatment		2,27	2,29	2,30	2,30	2,31		Passed	
23	Temperature conditioned	m .		2,21	2,22	2,25	2,27	2,26	FFP3 ≤ 3,0	Passed
24			2,26	2,28	2,30	2,31	2.30		Passed	
25			2,29	2,31	2,32	2,34	2.32		Passed	

Lab A



Page 10 / 12

UNIVERSAL SERTHIKASYON VE GÓZETIM HÍZM. TÍC. LTD. ŞTÍ. Keyap Ticaret Merkezi, Necip Fazal Bulván, E2 Blok, No. 44/84 V. Dudailta - Ümraniye - İSTANBUL, T+9/0 216 455 80 80 F+9/0 216 455 80 08 info@universalcert.com





# 7.17 CLOGGING (EN 149:2001 + A1:2009 clause 8.9, 8.10)

Test Method: Described in Clause 8.8, 8.10

REQUIREMENT	RESULTS	COMMENT
Valved particle filtering half masks: After clogging the inhalation resistances shall not exceed: FFP1:4mbar, FFP2:5mbar, FFP3:7mbar at 95L/min continuous flow. The exhalation resistance shall not exceed 3mbar at 160L/min continuous flow. Valveless particle filtering half masks: After clogging the inhalation resistances shall not exceed: FFP1:3mbar, FFP2:4mbar, FFP3:5mbar at 95L/min continuous flow	NAs	This is optional test and not desired by client.

Lab -

## 7.18 DEMOUNTABLE PARTS (EN 149:2001 + A1:2009 clause 8.2)

#### Test Method: Described in Clause 8.2

REQUIREMENT	RESULTS	COMMENT	-
All demountable parts (if fitted) shall be readily connected and secured, where possible by hand	N/A	No demountable part.	

Lab -

Pass	Requirement satisfied.
NCR	Requirement not satisfied. Refer to the "Result details" section for more information.
NAs Assessment not carried out.	
N/A	Requirement not applicable.

## LABORATORY INFORMATION

Code	Laboratory Name	Competency Explanations  Internal Laboratory Services of Notified Body		
Lab A	UNIVERSAL SERTIFIKASYON VE GOZETIM HIZMETLERI TIC. LTD. STI.			
Lab B	GCNTR ULUSLARARASI BELGELENDIRME, GOZETIM, EGITIM VE DIS TICARET LIMITED SIRKETI KOCAELI DILOVA SUBESI	Laboratory holds an accreditation by Turkish Accreditation Agency with number AB-1252-T according to EN ISO/IEC 17025:2017.		
	the laboratories is also under supervision / ass	NIVERSAL CERTIFICATION and the technical competence of sessment of UNIVERSAL CERTIFICATION based on the ats for bodies certifying products, processes and services standard, with the issuing laboratory code.		




Page 11 / 12

UNIVERSAL SERTIFIKASYON VE GÖZETİM HİZM. TİC. LID. ŞTİ. Keyap Ticaret Merkezi, Necip Farrl Bulvim, E2 Blob, No:44/84 Y. Dudullu - Ürnemiye - İSTANBUL T:+90 216 455 80 80 P:+90 216 455 80 08 info@universalcert.com





## Sample Photo



- End of Report -



Page 12 / 12

UNIVERSAL SERTIFIKAS YON VE GÖZETIM HIZM, TiC, LTD, ŞTİ, Keyap Ticaret Merkezi, Nocip Fazıl Bulvan, E2 Blok, No.44/84 Y. Dadaillu - Ümnuniye - İSTANBUL, T.+50 216 455 80 80 Ft+90 216 455 80 08 info@universalcert.com