PROMASK

Mascarillas FFP2 NR - Modelo Pez

Certificados y Declaración de Conformidad



PROMASK

mna LABORATUVARLARI
Notified Body Number: 2841

AB Tip İnceleme Sertifikası
EU Type-Examination Certificate

Belge No / *Certificate No* : 147-21-02-R01
Belgelendirme Tarihi - Bir Sonraki Belge Tarihi / *Certification Date / Certificate Validity Date* : 23.08.2021-04.02.2026
Belge Geçerlilik Tarihi / *Document Validity Period*: 5 yıl / *5 years*
Firma Unvanı ve Adresi / *Company Name and Address* : DESANTE ELEKTRONİK TEKNOLOJİ DIŞ TİC. A.Ş.
Beşyol Mah. Karadeniz Sk. Polaris Tekstil Blok No: 4 İçk: 4 Küçükçekmece / İSTANBUL

Ürün Adı /Modeller / *Product Name / Models* : ROSİMASK RM19930
Direktifi / *Directive* : 2016/425 REGULATION
Modülü/Kategori / *Module / Category* : B MODÜLÜ/ KATEGORİ III
MODULE B / CATEGORY III

Test Rapor No/ları / *Test Report No* : M-2021-00035
Ürün Tipi / *Product Type:*

- EN 149:2001+ A1:2009 Solunumla ilgili koruyucu cihazlar - Parçacıklara karşı koruma amaçlı filtreli yarım maskeler/ *Respiratory protective devices - Filtering half masks to protect against particles*

Ürünün Malzeme Bilgisi / ***Product Material Information***: ROSİMASK RM19930 model ürünleri kumaş, elastik kayış, burun klipsi ve filtre katmanı kullanılarak imal edilmiştir./ *ROSİMASK RM19930 model products are manufactured using fabric, elastic strap, nose clip, filter layer.*
Revizyon nedeni / ***Reason for revision***: Firma bilgileri revize edilmiştir./ *Company information has been revised.*

Volkan AKIN	*Okan AKEL*
23.08.2021	*23.08.2021*
Karar Verici / Approver	*Şirket Müdürü / General manager*

MNA Laboratuvarları San. Tic.Ltd .Şti
Adres: Küçükbakkalköy Mahallesi Yenidoğan Cad.No:21 Ataşehir/ İstanbul
Tel: 0216 574 07 08 Faks: 0216 575 13 31 www.mnalab.com

PROMASK

Notified Body Number: 2841

ATTACHMENTS (147-21-02-R01)

To certify the PPE product at Category III level, C2 or D module is accompanied by applying one of the conformity assessment methods along with the EU Type Examination (Module B).

Model : ROSİMASK RM19930

PPE SPECIFICATION	PERFORMANCE LEVELS
Classification	FFP2
Reusable / Single Shift Use	NR


PPE produced as a single unit to fit an individual user, all the necessary instructions for manufacturing such PPE on the basis of the approved basic model:

MARKING
MANUFACTURER: DESANTE ELEKTRONİK TEKNOLOJİ DIŞ TİC. A.Ş. **PPE TYPE:** - EN 149:2001+ A1:2009 Respiratory protective devices - Filtering half masks to protect against particles **MODEL:** ROSİMASK RM19930 **PICTOGRAM AND PERFORMANCE LEVELS:** EN 149:2001+ A1:2009 FFP2 NR  Or Condition of Storage

MNA LABORATORIES SAN. TIC. LTD. ŞTİ declares that the above-mentioned product meets the requirements of the directive according to the EU Directive 2016/425, the safety of the product is covered by the conditions and use specified in this certificate and in the technical file.

Notified Body Number: 2841

ATTACHMENTS (147-21-02-R01)

PRODUCT PICTURES

ROSİMASK RM19930

DOCUMENTS IN THE TECHNICAL FILE
- Basic Health Safety Requirements - Risk Assessment - Test Reports - Technical Report

MNA LABORATUVARLARI SAN. TİC. LTD. ŞTİ.	MNA LABORATUVARLARI
	TECHNICAL EVALUATION REPORT (147-21-02-R01)

Report No :147-21-02-R01

Report Date :23.08.2021


Application No :147-21-02-R01

1. **COMPANY INFORMATION:**
   DESANTE ELEKTRONİK TEKNOLOJİ DIŞ TİC. A.Ş.
   Beşyol Mah. Karadeniz Sk. Polaris Tekstil Blok No: 4 İçk: 4 Küçükçekmece / İSTANBUL

2. **PPE INFORMATION:**
   Disposable and non-sterile half mask made of particulate protection fitler material.

3. **PPE TYPE IDENTIFICATION**
   EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks to protect against particles - Requirements, testing, marking

4. **PPE PICTURES**



5. **PPE DIMENSIONS:**
   ROSİMASK RM19930 model has been found to be produced using standard sizes.

6. **PPE PRODUCT MATERIAL INFORMATION:**
   The product is made of elastic strap, nonwoven fabric on the outer and inner layers and fitler material on the middle layer.

7. ESSENTIAL HEALTH AND SAFETY REQUIREMENTS

- A visual inspection was made according to EN 149:2001 +A1:2009 for ergonomics.
- Protection levels and degrees are defined by the manufacturer.
- Suitable construction materials were determined by visual inspection according to EN 149:2001 +A1:2009.
- Respiratory protective dimensions are evaluated according to EN 149:2001 +A1:2009.
- Conditioning EN 149:2001 +A1:2009 part 8.3, Penetration EN 149:2001 +A1:2009 part 8.11 (EN 13274-7), Application performance EN 149:2001 +A1:2009 part 8.4, Inward leakage EN 149:2001 +A1:2009 part 8.5, Flammability EN 149:2001 +A1:2009 part 8.6, The carbon dioxide content of the inhaled air EN 149:2001 +A1:2009 part 8.7, Inhalation resistance EN 149:2001 +A1:2009 part 8.9, Exhalation resistance EN 149:2001 +A1:2009 part 8.9 has been tested and evaluated.

8. ANALYSIS AND EVALUATIONS:

EN 149:2001 +A1:2009

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Visual inspection	Shall also the marking and the information supplied by the manufacturer				Appropriate	-	PASS
Banned Azo Dyes	≤30 mg/kg				≤ 5 mg/kg	≤ 30 mg/kg	PASS
Total inward leakage	At least 46 out of the 50 individual exercise result	<25	<11	<5	See the table below	FFP2	PASS
	At least 8 out of the 10 individual wearer arithmetic means	<22	<8	<2	See the table below	FFP2	PASS

Total Inward Leakage (%)						
	Exercise 1	Exercise 2	Exercise 3	Exercise 4	Exercise 5	Average
Subject 1 (As recieved)	6.5	5.9	7.4	7.6	7.7	7.0
Subject 2 (As recieved)	6.7	5.9	4.7	7.8	7.6	6.5
Subject 3 (As recieved)	6.8	5.8	7.4	7.0	7.4	6.9
Subject 4 (As recieved)	6.5	8.0	7.3	8.7	7.6	7.6
Subject 5 (As recieved)	6.2	8.2	8.0	3.6	7.6	6.7
Subject 6 (After temperature conditioning)	6.1	6.8	6.6	6.1	7.6	6.6
Subject 7 (After temperature conditioning)	5.9	7.1	6.5	7.4	6.0	6.6
Subject 8 (After temperature conditioning)	6.2	6.5	6.2	7.6	7.5	6.8
Subject 9 (After temperature conditioning)	6.2	6.4	6.1	7.4	6.0	6.4
Subject 10 (After temperature conditioning)	6.3	7.4	7.6	7.4	6.2	7.0

MNA LABORATUVARLARI SAN. TİC. LTD. ŞTİ.	MNA LABORATUVARLARI
	TECHNICAL EVALUATION REPORT (147-21-02-R01)

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Flammibility	Mask shall not burn or not to continue to burn for more than 5 s				Flame not seen	-	PASS
Carbondioxide content of the inhalation air	Shall not exceed an average of % 1				0,72 0,79 0,71	-	PASS
Penetration of filter material	Sodium chloride, 95 L/min %, max	% 20	% 6	% 1	See the table below	FFP2	PASS
	Paraffin oil, 95 L/min %, max	% 20	% 6	% 1	See the table below	FFP2	PASS

Penetration of filter material	Sodium Chloride (%)	Paraffin Oil (%)
As recieved	3.7	4.0
As recieved	3.3	4.5
As recieved	4.1	4.1
After the simulated wearing treatment	3.2	3.8
After the simulated wearing treatment	3.3	4.2
After the simulated wearing treatment	3.0	4.6
Mechanical strength and temperature conditioning	4.9	5.5
Mechanical strength and temperature conditioning	5.1	5.1
Mechanical strength and temperature conditioning	5.3	5.4

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Compatibility with skin	Materials shall not be known to be likely to cause irritation or any other adverse effect to health				Appropriate	-	PASS
Head harness	It can be donned and removed easily				Appropriate	-	PASS
Breathing Resistance	Inhalation 30L/min	0,6 mbar	0,7 mbar	1 mbar	See the table below	FFP2	PASS
	Inhalation 95L/min	2,1 mbar	2,4 mbar	3 mbar	See the table below	FFP2	PASS
	Exhalation 160L/min	3 mbar	3 mbar	3 mbar	See the table below	FFP2	PASS

MNA LABORATUVARLARI SAN. TİC. LTD. ŞTİ.	MNA LABORATUVARLARI
	TECHNICAL EVALUATION REPORT (147-21-02-R01)

Breathing Resistance (mbar)	Inhalation 30L/min	Inhalation 95L/min
As recieved	0.3	2,0
As recieved	0.3	2,0
As recieved	0.3	2,1
After temperature conditioning	0.4	2,1
After temperature conditioning	0.3	2,1
After temperature conditioning	0.4	2,0
After the simulated wearing treatment	0.4	2,1
After the simulated wearing treatment	0.3	2,0
After the simulated wearing treatment	0.4	2,1

Breathing Resistance 160L/min (mbar)	Facing directly ahead	Facing vertically upwards	Facing vertically downwards	Lying on the left side	Lying on the right side
As recieved	2,1	2,2	2,1	2,1	2,2
As recieved	2,2	2,1	2,2	2,2	2,1
As recieved	2,1	2,1	2,1	2,2	2,2
After temperature conditioning	2,2	2,2	2,2	2,1	2,2
After temperature conditioning	2,1	2,2	2,1	2,2	2,2
After temperature conditioning	2,2	2,1	2,2	2,1	2,2
After the simulated wearing treatment	2,1	2,2	2,1	2,2	2,1
After the simulated wearing treatment	2,2	2,1	2,2	2,1	2,2
After the simulated wearing treatment	2,2	2,2	2,2	2,2	2,2

9. DECISION PROPOSAL

Analysis and examinations ROSİMASK RM19930 model coded personal protective equipment; Respiratory Protective Devices EN 149:2001 +A1:2009- Filtered Half Masks for Protection Against Particles - Properties, Experiments and Marking standards are evaluated. It is recommended to be certified at the performance levels specified as a result of technical evaluations.

10. ATTACHMENTS

- Basic Health Safety Requirements
- Risk Assessment
- Test Reports
- User Instruction

Reason for revison : Company information has been revised.

CONTROLLER : VOLKAN AKIN

SIGNATURE :

DATE : 23.08.2021

PROMASK

mna LABORATUVARLARI
Notified Body Number: 2841

CONFORMITY TO TYPE BASED ON INTERNAL PRODUCTION CONTROL PLUS SUPERVISED PRODUCT CHECKS AT RANDOM INTERVALS (MODULE C2)

MODÜL C2 - ÜRETİMİN DÂHİLÎ KONTROLÜ VE ÜRÜNÜN RASTGELE ARALIKLARLA DENETİMLİ MUAYENESİNE DAYALI TİPE UYGUNLUK

Belge No / *Certificate No*	: 147-21-02-01-R01
Belgelendirme Tarihi - Bir Sonraki Belge Tarihi / *Certification Date / Certificate Validity Date*	: 23.08.2021-12.07.2022
Belge Geçerlilik Tarihi / *Document Validity Period*:	1 yıl / *1 year*
Firma Unvanı ve Adresi / *Company Name and Address*	: DESANTE ELEKTRONİK TEKNOLOJİ DIŞ TİC. A.Ş. Beşyol Mah. Karadeniz Sk. Polaris Tekstil Blok No: 4 İçk: 4 Küçükçekmece / İSTANBUL
Ürün Adı /Modeller / *Product Name / Models*	: ROSİMASK RM19930
Direktifi / *Directive*	: 2016/425 REGULATION
Modülü/Kategori / *Module / Category*	: C2 MODÜLÜ/ KATEGORİ III *MODULE C2 / CATEGORY III*
Test Rapor No/ları / *Test Report No*	: M-2021-00553

Ürün Tipi / *Product Type:*

- EN 149:2001+ A1:2009 Solunumla ilgili koruyucu cihazlar - Parçacıklara karşı koruma amaçlı filtreli yarım maskeler/ *Respiratory protective devices - Filtering half masks to protect against particles*

Ürünün Malzeme Bilgisi / ***Product Material Information***: ROSİMASK RM19930 model ürünleri kumaş, elastik kayış, burun klipsi, filtre katmanı kullanılarak imal edilmiştir./ *ROSİMASK RM19930 model products are manufactured using fabric, elastic strap, nose clip, filter layer.*

Revizyon nedeni / ***Reason for revision***: Firma bilgileri revize edilmiştir./ *Company information has been revised.*

Erhan ÜSTÜNEL	*Okan AKEL*
23.08.2021	*23.08.2021*
Karar Verici / Approver	*Şirket Müdürü / General manager*

MNA Laboratuvarları San. Tic.Ltd .Şti
Adres: Küçükbakkalköy Mahallesi Yenidoğan Cad.No:21 Ataşehir/ İstanbul
Tel: 0216 574 07 08 Faks: 0216 575 13 31 www.mnalab.com

Notified Body Number: 2841

CONFORMITY TO TYPE BASED ON INTERNAL PRODUCTION CONTROL PLUS SUPERVISED PRODUCT CHECK AT RANDOM INTERVALS (MODULE C2, ANNEX VII) (147-21-02-01-R01)

Report No : 147-21-02-01-R01

Report Date : 23.08.2021


Application No : 147-21-02-01-R01

1. **COMPANY INFORMATION:**
   DESANTE ELEKTRONİK TEKNOLOJİ DIŞ TİC. A.Ş.
   Beşyol Mah. Karadeniz Sk. Polaris Tekstil Blok No: 4 İçk: 4 Küçükçekmece / İSTANBUL

2. **PPE INFORMATION:**
   Disposable and non-sterile half mask made of particulate protection fitler material.

3. **PPE TYPE IDENTIFICATION**
   EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks to protect against particles - Requirements, testing, marking

4. **PPE PICTURES**



ROSİMASK RM19930

5. **PPE DIMENSIONS:**
   ROSİMASK RM19930 model has been found to be produced using standard sizes.

6. **PPE PRODUCT MATERIAL INFORMATION:**
   The mask is made of elastic strap, nonwoven fabric on the outer and inner layers and fitler material on the middle layer.

Notified Body Number: 2841

CONFORMITY TO TYPE BASED ON INTERNAL PRODUCTION CONTROL PLUS SUPERVISED PRODUCT CHECK AT RANDOM INTERVALS (MODULE C2, ANNEX VII) (147-21-02-01-R01)

7. ESSENTIAL HEALTH AND SAFETY REQUIREMENTS

- A visual inspection was made according to EN 149:2001 +A1:2009 for ergonomics.
- Protection levels and degrees are defined by the manufacturer.
- Suitable construction materials were determined by visual inspection according to EN 149:2001 +A1:2009.

8. ANALYSIS AND EVALUATIONS:

EN 149:2001 +A1:2009

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Part 7.3 Visual inspection	Shall also the marking and the information supplied by the manufacturer				Appropriate	-	PASS
Banned Azo Dyes	< 30 mg/kg				< 5 mg/kg	-	PASS
Part 7.4 Packaging	Particle filtering half mask shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use.				Appropriate	-	PASS
Part 7.5 Material	When conditioned in accordance 8.3.1 & 8.3.2 the particle filter half mask shall not collapse.				Appropriate	-	PASS
Part 7.6 Cleaning and disinfecting	After cleaning and disinfecting the re-usable particle filtering half mask shall satisfy the penetration requirement of the relevant class.				Not applicable	-	Not applicable
Part 7.7 Practical performance	No negative comments should be made by the test subject regarding any of the criteria evaluated.				Appropriate	-	PASS
Part 7.8 Finish of parts	Parts of the device likely to come into contact with the wearer shall have no sharp edge or burrs.				Appropriate	-	PASS

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Part 7.9.1 Total inward leakage	At least 46 out of the 50 individual exercise result	<25	<11	<5	See the table below	FFP2	PASS
	At least 8 out of the 10 individual wearer arithmetic means	<22	<8	<2	See the table below	FFP2	PASS

Total Inward Leakage (%)

Notified Body Number: 2841

CONFORMITY TO TYPE BASED ON INTERNAL PRODUCTION CONTROL PLUS SUPERVISED PRODUCT CHECK AT RANDOM INTERVALS (MODULE C2, ANNEX VII) (147-21-02-01-R01)

	Exercise 1	Exercise 2	Exercise 3	Exercise 4	Exercise 5	Average
Subject 1 (As received)	7,0	6,0	6,0	7,2	5,5	6,3
Subject 2 (As received)	6,7	6,0	7,2	5,5	5,4	6,2
Subject 3 (As received)	6,4	6,5	4,9	7,2	7,6	6,5
Subject 4 (As received)	7,8	7,2	6,9	7,3	7,6	7,4
Subject 5 (As received)	6,1	7,3	6,7	4,4	6,2	6,1
Subject 6 (After temperature conditioning)	6,6	6,5	6,6	6,7	7,7	6,8
Subject 7 (After temperature conditioning)	6,4	6,6	6,3	5,3	6,2	6,2
Subject 8 (After temperature conditioning)	7,7	6,7	7,6	7,8	6,2	7,2
Subject 9 (After temperature conditioning)	6,0	6,5	7,6	7,2	6,5	6,8
Subject 10 (After temperature conditioning)	6,1	6,5	6,6	6,8	6,9	6,6

Subject facial dimensions

Subject	Face Length (mm)	Face Width (mm)	Face Depth (mm)	Mouth Width (mm)
1	133	132	132	65
2	125	144	116	67
3	126	135	124	75
4	123	133	134	74
5	117	135	122	73
6	122	142	133	66
7	113	132	114	75
8	135	123	123	65
9	122	135	133	74
10	135	142	125	83

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Part 7.9.2 Penetration of filter material	Sodium chloride, 95 L/min %, max	% 20	% 6	% 1	See the table below	FFP2	PASS
	Paraffin oil, 95 L/min %, max	% 20	% 6	% 1	See the table below	FFP2	PASS

Penetration of filter material	Sodium Chloride (%)	Paraffin Oil (%)
As received	4,4	4,4
As received	4,6	4,4
As received	4,4	4,8
After the simulated wearing treatment	4,3	5,0
After the simulated wearing treatment	4,7	4,7
After the simulated wearing treatment	4,6	5,1
Mechanical strength and temperature conditioning (120 mg)	5,6	5,7
Mechanical strength and temperature conditioning (120 mg)	5,8	5,9
Mechanical strength and temperature conditioning (120 mg)	5,6	5,8

TESTS	PARAMETER	PERFORMANCE	RESULTS	PERFORMANCE	EVALUATION

Notified Body Number: 2841

CONFORMITY TO TYPE BASED ON INTERNAL PRODUCTION CONTROL PLUS SUPERVISED PRODUCT CHECK AT RANDOM INTERVALS (MODULE C2, ANNEX VII) (147-21-02-01-R01)

		LEVELS				LEVELS	
		FFP1	FFP2	FFP3			
Part 7.10 Compatibility with skin	Materials shall not be known to be likely to cause irritation or any other adverse effect to health				Appropriate	-	PASS
Part 7.11 Flammibility	Mask shall not burn or not to continue to burn for more than 5 s				Flame not seen	-	PASS
Part 7.12 Carbondioxide content of the inhalation air	Shall not exceed an average of % 1				0,66 0,77 0,75	-	PASS
Part 7.13 Head harness	It can be donned and removed easily				Appropriate	-	PASS
Part 7.14 Field of vision	The field of vision shall acceptable in practical performance test.				Appropriate	-	PASS
Part 7.15 Exhalation valve(s)	It shall withstand axially a tensile force of 10 N apply for 10 s. If fitted, shall continue to operate correctly after a continuous exhalation flow of 300 L/min over a period of 30 s.				Not applicable	-	Not applicable

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Part 7.16 Breathing Resistance	Inhalation 30L/min	0,6 mbar	0,7 mbar	1,0 mbar	See the table below	FFP2	PASS
	Inhalation 95L/min	2,1 mbar	2,4 mbar	3,0 mbar	See the table below	FFP2	PASS
	Exhalation 160L/min	3,0 mbar	3,0 mbar	3,0 mbar	See the table below	FFP2	PASS

Breathing Resistance (mbar)	Inhalation 30L/min	Inhalation 95L/min
As received	0,5	1,7
As received	0,4	1,6
As received	0,4	1,6
After temperature conditioning	0,4	1,6
After temperature conditioning	0,5	1,6
After temperature conditioning	0,4	1,7
After the simulated wearing treatment	0,5	1,6
After the simulated wearing treatment	0,5	1,6
After the simulated wearing treatment	0,5	1,7
After the flow conditioning	-	-
After the flow conditioning	-	-
After the flow conditioning	-	-

Breathing Resistance 160L/min (mbar)	Facing	Facing	Facing	Lying on the	Lying on the

	directly ahead	vertically upwards	vertically downwards	left side	right side
As received	2,7	2,8	2,7	2,8	2,7
As received	2,8	2,8	2,7	2,7	2,8
As received	2,8	2,8	2,7	2,7	2,7
After temperature conditioning	2,8	2,7	2,8	2,7	2,8
After temperature conditioning	2,8	2,7	2,8	2,7	2,7
After temperature conditioning	2,8	2,7	2,8	2,7	2,7
After the simulated wearing treatment	2,8	2,7	2,8	2,7	2,8
After the simulated wearing treatment	2,8	2,7	2,8	2,7	2,8
After the simulated wearing treatment	2,7	2,8	2,7	2,8	2,7
After the flow conditioning	-	-	-	-	-
After the flow conditioning	-	-	-	-	-
After the flow conditioning	-	-	-	-	-

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Part 7.17 Clogging	After clogging the inhalation resistances shall not exceed. (valved)	4 mbar	5 mbar	7 mbar	Not applicable	-	Not applicable
	The exhalation resistance shall not exceed 3 mbar at 160 L/ min continuous flow. (valved)				Not applicable	-	Not applicable
	After clogging the inhalation and exhalation resistances shall not exceed. (valveless)	3 mbar	4 mbar	5 mbar	Not applicable	-	Not applicable
Part 7.18 Demountable part	All demountable parts (if fitted) shall be readily connected and secured were possible by hand.				Not applicable	-	Not applicable

9. DECISION

Analysis and examinations ROSİMASK RM19930 model coded personal protective equipment; Respiratory Protective Devices EN 149:2001 +A1:2009- Filtered Half Masks for Protection Against Particles - Properties, Experiments and Marking standards are evaluated. The homogeneity of the production was monitored at the performance levels determined as a result of the technical evaluations made within the scope of MODULE C2.

Notified Body Number: 2841

CONFORMITY TO TYPE BASED ON INTERNAL PRODUCTION CONTROL PLUS SUPERVISED PRODUCT CHECK AT RANDOM INTERVALS (MODULE C2, ANNEX VII) (147-21-02-01-R01)

10. ATTACHMENTS

- Basic Health Safety Requirements
- Risk Assessment
- Test Reports (M-2021-00553)
- User Instruction

Reason for revison	**: Company information has been revised.**
CONTROLLER	**: ERHAN ÜSTÜNEL**
SIGNATURE	**:**
DATE	**: 23.08.2021**

PROMASK

desante

EU-Konformitätserklärung

EU-Declaration of Conformity

Dokument-Nr. / *Document No.* 202008

Wir / *we* :Desante Elektonik Teknoloji Dış Ticaret A.Ş
Beşyol Mah. Karadeniz Sk. No:4 Kat:4 Küçükçekmece/İstanbul/ Türkiye

erklären in alleiniger Verantwortung, dass das Produkt
declare under our sole responsibility that the product

Partikelfiltrierende Halbmaske
Name/Typ: RM-19930 , Marke/Brand:ROSIMASK
(Größen: Standardgröße) (Sizes: Standard size)
Particle filtering half mask
Classification FFP2 NR

mit der EU-Baumusterprüfbescheinigung
is in conformity with the EU-Type Examination Certificate M-2021-00035

ausgestellt von der notifizierten Stelle mit der Kenn-Nr.
issued by the Notified Body with Identification No: 2841

Name und Anschrift der notifizierten Stelle
MNA Laboratuvarları san. Tic. Ltd. Şti.
Küçükbakkalköy Mahallesi Yenidoğan Caddesi No.21
Ataşehir/İst/Türkiye

und mit der folgenden Harmonisierungsrechtsvorschrift der Union unter Anwendung der aufgeführten Norm übereinstimmt:
and is in compliance with the following Union harmonisation legislation by application of the listed standard

Bestimmungen der Verordnung *provisions of regulation*		**Nummer sowie Ausgabedatum der Norm** *Number and date of issue of standard*
Verordnung (EU) 2016/425	**Verordnung über persönliche Schutzausrüstungen** *Personal Protective Equipment Regulation*	EN 149:2001+A1:2009

o oder: Überwachung der Konformität mit der Bauart auf der Grundlage einer internen Fertigungskontrolle mit überwachten Produktprüfungen in unregelmäßigen Abständen (Modul C2) durch:
Surveillance of conformity to type based on internal production control plus supervised product checks at random intervals (Module C2)
by:MNA Laboratuvarları san. Tic. Ltd. Şti.
Küçükbakkalköy Mahallesi Yenidoğan Caddesi No.21 Ataşehir/İst/Türkiye
Zertifikat-Nr. *147-21-03*
Certificate No.

İstanbul/ 24.08.2021
Ort und Datum
Place and date

Cevdet Darende
General Manager
Name/Funktion
Name/Function

DESANTE ELEKTRONİK TEKNOLOJİ DIŞ TİCARET A.Ş.
Beşyol Mah. Karadeniz Sok. No: 4 Kat :4
Küçükçekmece/İSTANBUL
Küçükçekmece V.D. : 362 [illegible] 20 00
Ticaret No: 249865-5