



Mascarilla quirúrgica

Mascarilla filtrante de 3 capas no reutilizable.

Sin válvula. MEDICA.

Filtración mayor al 98% de las partículas.



Documentación técnica

- Ficha técnica
- Declaración de conformidad
- Test report
- Test report material

Mascarilla quirúrgica

Mascarilla quirúrgica de 3 capas. No reutilizable. MÉDICA. Filtración mayor al 98% de las partículas.



COLORES:

- Blanco
- Azul
- Rosa
- Naranja

COMPOSICIÓN:

Mascarilla de 3 capas compuestas de tejido no tejido "Melt Blown" (Polipropileno).
Gomas elásticas de poliéster.

ESPECIFICACIONES:

- Eficiencia de filtrado: BFE > 98%.
- Mascarilla Tipo II.
- No estéril. Filtrado microbiológico.
- No reutilizable.
- Vida útil: 5 años.
- Talla: adultos.
- Dimensiones: 176 x 95 mm.
- Gomas suaves y elásticas para un ajuste cómodo en ambos pabellones auditivos.
- No irrita la piel y no causa alergias.



3 capas de protección



Transpirable

PACKAGING:

- Cajas pequeñas: 50 unidades.
- Cajas distribución: 20 cajas pequeñas, 1000 unidades de mascarillas.
Dimensiones caja distribución: 52x38x40 cm
Peso bruto caja distribución: 7,6 kg.

NORMAS Y CERTIFICACIONES:

Este dispositivo es un producto sanitario según la Directiva 93/42 o del Reglamento UE / 2017/745.



Certificado Europeo CE
Nº de informe: 555-15-20-1
UNE EN-14683

Mascarilla certificada en LOTRIC, centro autorizado
localizado en Eslovenia.
Número de organismo notificado: 1304.



LOTRIC Metrology Ltd.

Dirección: Selca, 163 - 4227 Selca - Eslovenia
Teléfono: +386 4 51 70 700
E-mail: info@lotric.si
Página web: www.lotric.si

FOTOGRAFÍAS DEL PRODUCTO:

Caja de 50 unidades

**INSTRUCCIONES DE USO:**

1. Lavarse las manos durante 40-60 segundos antes de manipularla.
2. Tocar solo las gomas de la mascarilla.
3. Colocar la mascarilla sobre la nariz y boca.
4. Pasar las gomas elásticas por detrás de las orejas.
5. Pellizcar la banda metálica para ajustarla a la nariz.
6. Evitar tocar la parte exterior de la mascarilla. Si ocurre, lavar las manos antes y después.
7. Lavar las manos antes de retirarse la mascarilla.
8. Retirarla tocando solo las gomas elásticas.
9. Para desecharla, introducirla en una bolsa de plástico. Depositarla cerrada en la basura y lavar las manos.

Condiciones de almacenaje: Almacenar en un ambiente seco y fresco, alejado de la luz directa del sol y fuentes de calor.

Mascarilla quirúrgica



PROIZVODNJA MEDICINSKIH, STOMATOLOŠKIH I ZAŠTITNIH PROIZVODA ZA JEDNOKRATNU UPOTREBU
9. SEPTEMBAR MEDICAL DOO PIB: 108547827 TELEFON: +381 32 700 200
 32300 Gornji Milanovac MATIČNI BROJ: 21021652 E-MAIL: office@9smedical.com
 Velerečki put bb, Srbija ŽIRO RAČUN: 155-26653-76 170-30021251000-17 www.9smedical.com

EC DECLARATION OF CONFORMITY

Declaration No. **04/2020**

Manufacturer: "9. SEPTEMBAR MEDICAL" d.o.o.
 Velereč bb, 32300 Gornji Milanovac, SERBIA

Product Name: **Surgical Mask**

Model/Type: Three-layer Surgical masks with ear loop or with binding ribbon; white, green, blue, pink or orange color

Classification: Class I (*Directive 93/42/EEC of medical devices Annex IX*)

We "9. SEPTEMBAR MEDICAL" d.o.o. declare under our sole responsibility that the following products are fully complying with the Essential Requirements of Directive 93/42/EEC of Medical devices, Annex VII according to section 5 of article 11.

Standards applied:

- ISO 9001:2015 Quality management systems, Requirements
- EN ISO 13485:2016 Medical devices - Quality management systems – Requirements for regulatory purposes
- EN ISO 14971:2012
- ISO 10993 – Biocompatibility of products – Medical devices
- EN ISO 14683:2019 Bacteriological filtration efficiency of surgical masks - Differential respiratory pressure
- SRPS EN ISO 15223-1:2017 - Symbols used to mark a medical device and labeling packaging
- Ph.Jug V/2000 - Microbiological safety of products – Bioburden (cfu/g)

Notified body: SIQ, Slovenian Institute of Quality and Metrology
 Tržaška cesta 2
 1000 Ljubljana, SLOVENIA
 Notified Body No. 1304



Authorized representative in EU:

Medical Device Safety Service GmbH (MDSS)
 Schiffgraben 41, D-30175
 Hannover, Germany





PROIZVOĐAJA MEDICINSKIH, STOMATOLOŠKIH I ZAŠTITNIH PROIZVODA ZA JEDNOKRATNU UPOTREBU

9. SEPTEMBAR MEDICAL DOO

PIB: 108547827

TELEFON: +381 32 700 200

MATIČNI BROJ: 21021652

E-MAIL: office@9smedical.com

32300 Gornji Milanovac
Velerački put bb, Srbija

ŽIRO RAČUN: 155-26653-76

170-30021251000-17

www.9smedical.com

G. Milanovac, April 23th 2020.

Responsible person:

Zoran Jokanović, director

Zoran
Jokanović
364038-091
0962710200

Digitally signed

by Zoran
Jokanović

364038-0910962

710200

Date: 2020.08.17

14:59:40 +02'00'



Mascarilla quirúrgica

LOTRIČ[®] METROLOGY

Številka certifikata
certificate number

555-15-20-1

CERTIFIKAT O PRESKUSU

TEST CERTIFICATE

naročnik applicant "9. SEPTEMBAR MEDICAL" d.o.o.
Velereč bb, 32300 Gornji Milanovac

lastnik owner "9. SEPTEMBAR MEDICAL" d.o.o.
Velereč bb, 32300 Gornji Milanovac

vzorec sample Osebna polobrazna maska
Personal face halfmask
proizvajalec manufacturer 9. SEPTEMBAR MEDICAL

Identifikacijska številka
identification number
1-5 (SAMPLE 1)

LOT številka
LOT number
374/20

tip maske
mask type
II

vzorčenje sampling izvedel naročnik
done by applicant
Podrobnosti so podane v poglavju vzorčenje.
Details are given in chapter sampling.
rezultat result **ustreza** *meets*
Podrobnosti so podane v poglavju merilni rezultati.
Details are given in chapter measurement results.
datum preskusa
date of test
15.06.2020

datum odobritve
date of approval
19.06.2020

izvedel performed by
Žan Kavčič
Internally digitally signed


odobril approved by
Primož Hafner
odgovorna oseba
responsible person
digitally signed
date: 19.06.2020



Ta dokument se sme objavljati ali posredovati le v celoti. Verodostojnost podpisa se lahko preveri v elektronski obliki dokumenta.
This document may be published or forwarded only in full. Signature validity can be verified in electronic version.

LOTRIČ Meroslavje d.o.o., Selca 163, 4227 Selca
T +386 4 517 07 00, F +386 4 517 07 07, E info@lotric.si, W www.lotric.si

stran page
1/4

LOTRIČ[®] METROLOGY

Številka certifikata
certificate number

555-15-20-1

vzorčenje sampling

Vzorčenje je opravil naročnik, ki je poslal označene vzorce v laboratorij.
Sampling was performed by the applicant who sent the samples to the laboratory.

mesto vzorčenja
place of sampling

"9. SEPTEMBAR MEDICAL" d.o.o.
Velereč bb, 32300 Gornji Milanovac

preskusni postopek
test procedure

Preskus je bil izveden po navodilu ML10N211 - z merjenjem padca tlaka pri stalnem pretoku 8 l/min na aparatu za določanje prepustnosti zraka, s predhodno pripravljenimi vzorci v skladu z evropskim standardom EN 14683:2019+AC:2019 (točka 5.2.3 in aneks C). Preskus se je opravil na sredini vzorca s presekom 20 cm², s tem da se pri nastavitvah aparata upoštevata zahteva standarda za preskus na preseku 4,9 cm².

The test was carried out following the instruction ML10N211 - by measuring the differential pressure at constant flow rate of 8 l/min on an air permeability tester, with conditioned samples in accordance with European standard EN 14683:2019+AC:2019 (section 5.2.3 and annex C). The test was carried out in the center of the sample total area of 20 cm², considering the requirement of the standard for a 4,9 cm² total area when adjusting the tester.

mesto preskusa
place of test

FILC d.o.o.
Laboratorij – mehanski del, Trata 48, 4220 Škofja Loka

pogoji okolja
environmental conditions

		od from	do to	dovoljeno odstopanje v času meritev tolerance during measurements
temperatura zraka air temperature	(°C)	23,4	23,4	± 1
relativna vlaga relative humidity	(%)	52,2	52,2	± 5

št. certifikata uporabljenih referenčnih etalonov

certificate no. of
reference standards used

251-250-19-1, KB21403-18-406.02.15

sledljivost traceability

Pri izvajanju meritev so bili uporabljeni etaloni, ki so sledljivi do nacionalnih etalonov in s tem do mednarodno podprtih realizacij SI-enot. Sledljivost je zagotovljena s kalibracijo v ustreznem kalibracijskem laboratoriju.

The reported measurement values are traceable to the national measurement standards and thus to internationally supported realizations of the SI-units. Traceability is ensured by calibration in the relevant calibration laboratory.

stanje vzorca pri prevzemu (prejeto stanje)

state of sample at acceptance
(as found)

Vzorci so bili shranjeni v izvorni embalaži in brez vidnih poškodb.
The samples were stored in original packaging and without visible damage.

Velikost vzorca je / Sample size is: 175-180 mm x 90-95 mm (podatek proizvajalca / manufacturer's information)

 slike vzorca
sample pictures

vzorci pred preskusom (prejeto stanje)
samples prior to the test (as found)



vzorci po preskusu (končno stanje)
samples after the test (as left)



postopek predpriprave
 conditioning procedure

Predpriprava je bila izvedena po navodilu ML10N211 – v klimatski komori v trajanju najmanj 4 ure pri temperaturi $21\text{ °C} \pm 1\text{ °C}$ in relativno vlago $85\% \pm 2,5\%$ v skladu z evropskim standardom EN 14683:2019 (aneks C).

The conditioning was carried out following the instruction ML10N211 – in the climatic chamber for at least 4 hours at temperature $21\text{ °C} \pm 1\text{ °C}$ and relative humidity $85\% \pm 2,5\%$ in accordance with European standard EN 14683:2019 (annex C).

merilni rezultati
 measurement results

številka vzorca sample number	izmerjeni padec tlaka measured differential pressure (Pa/cm ²)	merilna negotovost uncertainty (Pa/cm ²)	zahteva requirement (Pa/cm ²)	
1	13,2	0,5	< 40	✓
2	18,1	0,7	< 40	✓
3	16,2	0,6	< 40	✓
4	13,3	0,5	< 40	✓
5	13,9	0,5	< 40	✓
povprečje average	14,9	AQL 4%	< 40	✓

LOTRIČ[®] METROLOGY

Številka certifikata
certificate number

555-15-20-1

izjava o skladnosti
statement of compliance

Zahteva je določena glede na evropski standard EN 14683:2019, za padec tlaka (točka 5.2.7).
Requirement is determinate according to European standard EN 14683:2019 for pressure drop (section 5.2.7).

- ✓ Merilni rezultati so ZNOTRAJ zahtev.
The measurement results are WITHIN the requirements.
- ✗ Merilni rezultati so ZUNAJ zahtev.
The measurement results are OUTSIDE the requirements.

Izjava o skladnosti je podana brez upoštevanja razširjene merilne negotovosti.
Statement of compliance is based without considering expanded uncertainty.

izjava statement

Podani merilni rezultati se nanašajo izključno na preskušane vzorce in izmerjene vrednosti v času meritev, ki ne zagotavljajo dolgotrajne stabilnosti ali razširjanje rezultatov na ostale vzorce ali serije vzorcev.
The measurement results refer only to the tested item and to the measured values at the time of measurement, which carry no implication regarding the long term stability or disseminating results to other samples or series of samples.

n/a – se ne uporablja
n/a - not applicable

merilna negotovost
uncertainty

Podana razširjena merilna negotovost je podana kot standardna negotovost pomnožena s faktorjem pokritja $k = 2$, kar za normalno porazdelitev ustreza intervalu verjetnosti približno 95 %. Standardno negotovost smo določili v skladu z EA vodilom EA-4/02 M: 2013.
The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95 %. The standard uncertainty of measurement has been determined in accordance with EA Publication EA-4/02 M: 2013.

Mascarilla quirúrgica



**Innovatec Microfibre Technology GmbH & Co.
KG
Mrs. Daniela Schlösser
Gierlichsstraße 10
53840 TROISDORF
Allemagne**

Your notice of
02-04-2020

Your reference

Date
24-04-2020

Analysis Report 20.01913.01

Required tests :

EN 14683 (2019) + AC
(2019)

EN 14683 - annex B (2019)
+ AC (2019)

Bacterial filtration efficiency

EN 14683 (2019) + AC
(2019)

EN 14683 - annex C (2019)
+ AC (2019)

Medical face masks - Breathability
(differential pressure)

Identification number	Information given by the client	Date of receipt
T2007278	1F1020C00N	02-04-2020


Sylvie Niessen
Order responsible

This report may be reproduced, as long as it is presented in its entire form, without written permission of Centexbel.
The results of the analysis cover the received samples. Centexbel is not responsible for the representativeness of the samples.
In assessing compliance with the specifications, we did not take into account the uncertainty on the test results.

RIJCHTVOG ERKENNEN TOEPASSING VAN DE RESULTAAT VAN 30 JANUARI 1947 / ETABLISSMENT RECONNU PAR APPLICATION DE L'ARRÊTÉ-CK DU 30 JANVIER 1947



Analysis Report 20.01913.01
Date 24-04-2020
Page 2/5

Reference: T2007278 - 1F1020C00N

Bacterial filtration efficiency

Date of ending the test	14-04-2020
Standard used	EN 14683 - annex B (2019) + AC (2019)
Product standard	EN 14683 (2019) + AC (2019)
Sample description :	Meltblown filter layer for medical face masks - One layer - White - 20g/m ²
Number of tested specimens :	3
BFE Area tested :	± 49 cm ²
Test specimen conditioning :	21 ± 5°C and 85 ± 5% RH
Side of the specimen in contact with the bacterial challenge :	Inner side
Challenge bacterial strain used :	<i>Staphylococcus aureus</i> ATCC6538
Bacterial challenge per test :	1700 - 3000 CFU
Total test time :	1 min. delivering challenge + 1 min. without challenge (air flow continuing)
Flow rate :	28.3 l/min.
Positive control	Tests performed with no filter material in the air stream
Negative control	Test performed without challenge
Deviation from the standard	The test has not been performed on a mask but only on one layer of the tested material Test result based on 3 instead of 5 samples

IMPLICHTING ERKENNIS BIJ TOEGANGS VAN DE BESLUIT VAN 30 JANUARI 1947 / ETABLISSEMENT RECOGNU PAR APPLICATION DE LA REJETE-OF 30 JANUARI 1947

Performed in the microbiological lab under the responsibility of Yvette Register



Analysis Report 20.01913.01
Date 24-04-2020
Page 3/5

Results

B = Bacterial filtration efficiency (%)

$$B = \frac{(C - T)}{C} \times 100$$

With C = mean of the total plate counts for the positive control runs
T = total count for the tested specimen

# test specimen	B (%)
1	99,0
2	98,4
3	97,9

Mean particle size of the bacterial challenge aerosol : 2.8 µm


Controls

Mean positive controls 2290 CFU
Negative control < 1 CFU

This test report is valid for products used in relation to the current Covid-19 health crisis and for products which are not entering the regular distribution channels. Cfr Commission Recommendation (EU) 2020/403 of 13 March 2020 on conformity assessment and market surveillance procedures within the context of the COVID-19 threat"

IMBICHT 'NG EIRWNG BIJ TOEPASSING VAN DE BESLUITING VAN 30 JANUARIJ 1947 / ETABLISSEMENT RECONNU PAR APPLICATION DE L'ARRÊTÉ DU 30 JANVIER 1947

Performed in the microbiological lab under the responsibility of Yvette Rogister




Analysis Report 20.01913.01
Date 24-04-2020
Page 4/5

Reference: T2007278 - 1F1020C00N

Medical face masks - Breathability (differential pressure)

Date of ending the test	05-04-2020
Standard used	EN 14683 - annex C (2019) + AC (2019)
Product standard	EN 14683 (2019) + AC (2019)
Sample description :	Meltblown filter layer for medical face masks - One layer - White - 20g/m ²
Number of tested specimens :	5
Number of areas per specimen	5 (see figure)
Dimension of the areas :	Disc whose diameter is 2.5 cm
Surface areas :	4,9 cm ²
Flow rate :	8 l/min.
Direction of the air flow :	From the inside of the mask to the outside
Test specimen conditioning :	21 ± 5°C and 85 ± 5% RH
Deviation from the standard	The test has not been performed on a mask but on one layer of the tested material

Figure : Distribution of the areas in the test specimen



RIJICHTING ERKENNEN BIJ TOEPASSING VAN DE BESLUITWIJZING VAN 30 JANUARIJ 1947 / ETABLISSEMENT RECOGNU PAR APPLICATION DE L'ARRÊTÉ-LOI DU 30 JANVIER 1947

Performed in the microbiological lab under the responsibility of Yvette Rogister



Analysis Report 20.01913.01
Date 24-04-2020
Page 5/5

ONMICHTIG ERKENNEN VAN DE BESLUITWIJZING VAN 30 JANUARIJ 1947 / ÉTABLISSEMENT RECONNU PAR APPLICATION DE L'ARRÊTÉ-LOI DU 30 JANVIER 1947

Results ΔP

	Test specimen 1	Test specimen 2	Test specimen 3	Test specimen 4	Test specimen 5
Area A	15.7	18.1	18.1	13.6	14.7
Area B	17.9	18.3	15.1	16.1	14.9
Area C	17.5	19.1	14.5	19.4	16.1
Area D	17.9	16.9	16.1	16.1	14.7
Area E	17.5	17.7	18.7	16.9	16.1
Average ΔP (Pa/cm²)	17.3	18.0	16.5	16.4	15.3

Performed in the microbiological lab under the responsibility of Yvette Rogister

Mascarilla quirúrgica



MDSS
Medical Device Safety Service

Certificate of CE-Registration

This is to certify that, in accordance with the Medical Device Directive 93/42/EEC. Medical Device Safety Service GmbH (MDSS) agrees to perform all duties and responsibilities as the Authorized Representative for:

**9. Septembar Medical d.o.o.
Veleroč bb
32300 GORNJI MILANOVAC
REPUBLIC OF SERBIA**

as stipulated and demanded by the aforementioned Directive. The German Competent Authority has allocated the medical devices of the Manufacturer registration numbers as foreseen in:

Annex A dated March 06, 2015

The Manufacturer has provided MDSS with the appropriate Declaration(s) of Conformity confirming that the medical devices fulfill the applicable requirements of Directive 93/42/EEC. In compliance with German law, a safety officer has been appointed for Germany.

2015-03-06


Dr. Huan Sun
Senior Manager Registration
MDSS GmbH

MDSS · Medical Device Safety Service · Schiffgraben 41 · 30175 Hannover, Germany